



Follow up on Advisory Notice issued in March 2021

*Subset of Assurity™ AND Endurity™ pacemakers
Models: PM1152, PM1160, PM1172, PM1240, PM1272, PM2152,
PM2160, PM2172, PM2240, PM2260, PM2272
Intermittent incomplete mixing of epoxy during manufacture*

DISTRIBUTOR ACKNOWLEDGEMENT TEMPLATE

I hereby acknowledge receipt of the instructions and associated materials necessary for the proper and timely implementation of this field action.

We will translate the Field Safety Notice in following language versions:

.....

We are aware that the translated version should be as close as possible to the original English text. We will not change the date stated in the English version of letter.

Distributor name:
Country:
Date (DD/MM/YY)
Signature:

Please fill out this form on receipt and return to your RA contact



В ответ на направленное в марте 2021 года уведомление с рекомендациями

*Ряд кардиостимуляторов Assurity™ и Endurity™
Модели: PM1152, PM1160, PM1172, PM1240, PM1272, PM2152,
PM2160, PM2172, PM2240, PM2260, PM2272
Прерывистое неполное смешивание эпоксидной смолы во время
производства*

ШАБЛОН ФОРМЫ ПОДТВЕРЖДЕНИЯ ДИСТРИБЬЮТОРА

Я настоящим подтверждаю получение мною инструкций и связанных с ними материалов для надлежащего и своевременного принятия указанного действия на местах.

Мы переведем Уведомление о действиях на местах на следующие языки:

Мы осведомлены о том, что перевод должен быть максимально близок к оригинальному тексту на английском языке. Мы не будем менять дату, указанную в письме на английском языке.

Наименование Дистрибьютора:
Страна:
Дата (дд.мм.гггг)
Подпись:

Заполните настоящую форму по получении и верните её вашему контактному лицу RA.